Hitachi High-Technologies

Hitachi Fluorescence Spectrophotometer

F-7000



19/11/

F-7000





Designed to Meet Your Needs for High-Quality Analytical Instrumentation.

Hitachi's Superior Fluorescence Technology Has Created a New Generation of Fluorescence Spectrophotometers.

High S/N Ratio, Ultra-Fast Scanning, Compact Design, Multiple Accessories

HITACH

F-7000

High sensitivity measurement (S/N 800 RMS) (equivalent to S/N 250 peak to peak)

60,000 nm/min ultra-high scanning speed, ideal for 3-D measurement.

Compact design (approx. 2/3 the size of the F-4500)

A wide range of accessories accommodating various applications

000

Outstanding Performance supported by Superb Manufacturing Technologies

> Sensitivity, Scan Speed, 3-D Measurement, and more....

> > Easy-to-Use Software



Technologies Supporting Hitachi Fluorescence Spectrophotometers

Precision Machining Technology resulting in bright optics.

Advanced Electric Circuit Technology for high-speed processing.

Controlled System Technology ensures high accuracy.



Stigmatic concave diffraction grating, mechanically ruled, resulting in a very bright monochromator of F-number 2.2.

Ruling engine.

A dividing engine for ruling diffraction gratings, invented in 1880s by Henry Augustus Rowland of Johns Hopkins University. Compared to a holographic grating, mechanically ruled gratings have

- the following advantages:
- (1) Mirror-finished groove surface results in high diffraction efficiency. (2) Groove spacing required for aberration correction can be adjusted, making it possible to have a greater correction effect.
- These characteristics of mechanically ruled gratings work well to create an excellent monochromator.

F-7000 Performance Supported by Technology

High S/N: 800 (RMS), 250 (Peak to Peak)

Detection limit of fluorescein

Due to enhanced sensitivity, the F-7000 Fluorescence Spectrophotometer can detect fluorescein concentrations one digit lower than its predecessor, the Model F-4500. The improved S/N ratio provides greater capabilities for trace-sample measurements.



3-D Measurement

Data of fluorescent marker pen

A 3-dimensional fluorescence spectrum can clearly distinguish slight differences that a 2-dimenstional spectrum cannot detect.

Measurements can now be carried out with higher accuracy than before.



-dimensional spectrum of fluorescent marker pen $(areen \times red)$



2-dimensional spectrum (excitation wavelength 460nm) fluorescent marker pen (green)

High-Speed Scan at 60.000nm/min

Example of reaction tracing with a spectrum

The fast-speed scanning enables users to carry out measurements that have been difficult with conventional instruments. In this example, an isomerization process of coumarin in kerosene was traced by spectrum measurement at 2-second intervals and displayed as a 3-D timescan spectrum. This is a new function in the F-7000. Previously, a guick reaction which occurs within 1 minute could be measured only by using the fixed-wavelength method. The F-7000 is capable of following such a guick reaction because of its fast scanning, measuring the entire wavelength range within 1 second.

Measures up to 6-digit concentration values

Calibration curve and spectrum overlay of fluorescein

The automatic gain change-over function, a technique unique to Hitachi fluorescence spectrophotometers, has made it possible to generate calibration curves using up to 6-digit concentration values. An unknown sample can be quantitatively analyzed without additional sample preparation.



Other functions

- Automatic pre-scan function optimized for unknown sam-• High-resolution multi-stage slit with a resolution as small ple measurement as 1nm
- Ratio photometry (0 point correction) ensuring stable measurements





3-dimensional time scan spectrum of coumarin for environmental applications

• Shutter control for minimizing sample deterioration

Application Capabilities Unique to Hitachi

NAMES OF TAXABLE PARTY.

Scan speed: 12,000nm/min

Photomultiplier voltage: 400V

Beam-cut filter (UV-39) used

Photomultiplier R928F used

Excitation slit: 5.0nm

Emission slit: 5.0nm

Response: Automatic

CONTRACTOR NO.

O STATEMENT DESCRIPTION OF THE OWNER OWNE COLUMN STREET TALKSTRASS CO.

Industrial Material Field

Measurement of fluorescent materials

Organic EL material

In this example, the F-7000 was used to analyze the luminescent characteristic of trisaluminum complex powder used as a luminescent material for organic EL display. A solid sample holder, its powder cell, the photomultiplier R928F, and the filter set were used.



The acquisition of these data was made possible by the 3-D measurement function and high-speed scanning capability of the F-7000.

Pharmaceutical Field

Phosphorescence measurement

Rare earth element complex (Eu chelate)

The example below shows the phosphorescence spectrum and lifetime measurement of the Eu(tta)3 (TOPO)2 complex, a rare earth element.



Phosphorescence spectrum measurement of Eu (tta)₃(TOPO)₂ complex

With the F-7000, the analysis of phosphorescence life of 1 ms order can be performed at room temperature without special accessories.



Phosphorescence life measurement of Eu(tta)₃(TOPO)₂ complex

Measurement of intermolecular actions

■ FRET (Fluorescence Resonance Energy Transfer) and BRET (Bioluminescence Resonance Energy Transfer)

The Model F-7000 can measure the intermolecular activities such as FRET and BRET. Shown below are fluorescence spectra presenting the interactions between the sub-



Measurement of calcium in cell

■ Ca²⁺ concentration in cells

With the optional interacellular calcium measurement accessory, the F-7000 can measure fluorescence intensity values at two wavelengths in EGF-injected COS-7 cells (extracted from a monkey's kidney), and calculate the concentrations of Ca²⁺.

The sample was a cultivated cell fluorescence-labeled by Fura2-AM.

The change in Ca²⁺ concentrations in the live cell was also measured. During this analysis, the EGF receptor appeared in the COS-7 as the Ca2+ level increased due to EGF injection.

The Model F-7000 can measure biological samples with higher sensitivity and speed.

Micro-plate Accessory

Features

- Automatic measurement using a 96 well microplate is available.
- Used as an autosampler, allows wavelength scan, time scan, and 3-dimensional measurement.
- Can be used in conjunction with the polarization accessory.

5



Biological Field

- unit proteins of an ATP-active purine receptor.
- Data provided by Mr. Takaaki Koshimizu, Kyoto University Graduate School of Pharmaceutical Sciences - Genomic Drug Discovery Science.





Easy-to-Use Software with Powerful Functionality

CONTRACTOR OF STREET

a walk of the second

7

The FL Solutions Software is a powerful tool for analysts to use a Hitachi F-7000 fluorescence spectrophotometer efficiently at their command and thereby generate the necessary reports.







New functions

■ Collective file conversion Multiple files can be converted simultaneously.











一 一 二 二 二 一 二 二 二 二 二 二 二 二 二 二 A Luxurious Array of Accessories for Applications in Extensive Fields

FLOW CELL UNIT FOR 55uL (250-0331)

FLOW CELL UNIT FOR 180uL (250-0332)

Provides high sensitivity measurements due to a design that avoids measuring fluorescence near the flow path.

An increased cell capacity is particularly effective for high sensitivity analysis of elements such as catecholamines when measured in combination with a high performance liquid chromatography system.





High sensitivity cell holder (650-0184)

Enhances sensitivity about two fold when used with the 10-mm rectangular cell. Compatible with the 10-mm rectangular cell (not included)

Compatible cell 10mm rectangular cell (Cell must be prepared separately.)



Sample sipper accessory (5J0-0123)

Streamlines successive operations of sample sipping, measurement and result printout. Effective for automatic measurement of liquid samples in quality control and clinical chemical analysis.





■ Thermostatic cell holder (250-0330)

THE R. S. LEWIS CO., LANSING MICH.

Temperature-controlled water keeps the temperature of the 10-mm rectangular cell constant. This holder is appropriate for analysis of biochemical samples.

Temperature range 5 to 60°C (Requires, but does not includes a thermostatted water bath and a cell)



4-turret sample compartment (250-0339)

Ideal for quantitative analysis when using 10mm rectangular cells.

Max. error due to cell changeover 3%, with the same sample and cell (Cell must be prepared separately.)



8-turret sample compartment (250-0333)

Effective for multi-sample measurements. Allows selection of up to eight 10mm rectangular cells/test tubes for rapid quantitative analysis.

Compatible cells 10mm rectangular cell Test tube of outer dia. 10/12mm and height 105mm or less Error due to cell changeover Max. 3% in signal level difference with the same sample and 10mm rectangular cell

(Cell not included)



Low temperature measurement accessory (5J0-0112)

Used for fluorescence/phosphorescence measurement at a liquid-nitrogen temperature. The micro-structure of a sample which does not appear at normal temperature can be measured with this accessory.

Sample tube	Outer dia. 5 or 8 mm	
Measurement temperature	–196°C	
iquid pitragan temporatura)		



Absorbance Cell holder (650-0165)

Used for measuring absorbance.

Allows to measure absorbance without influence from fluorescence due to the simultaneous scanning using the excitation and emission wavelengths (in synchronous spectrum measurement mode).

Compatible with the 10-mm rectangular cell (not included)



Thermostatic cell holder with stirrer (250-0346)

A magnetic stirrer is used to stir sample solutions to ensure higher thermal accuracy in measurement.

Minimum sample requirement	2.5mL (10mm rectangular cell), 0.4mL (micro-flow cell)
Stirrer speed	500 to 1,200rpm
Temperature range	5 to 60°C

Thermostatted water bath and cell required, but not included.



Micro sampling assembly (5J0-0111)

Used in combination with the thermostatted cell holder with stirrer (P/N 250-0346). A reagent can be injected using a micro syringe, without opening the sample compartment. Facilitates the measurement of a reaction

process after injecting a reagent. (Micro syringe required, but not included.)

Intracellular Cation measurement program (5<u>J0-0308</u>)

Software for measuring calcium (Ca) in cells. Can be used with pH measurement reagent (such as BCECF) along with Ca measurement reagents (Quin 2, Fura 2, Indo 1). Up to 4 sets of measurement wavelengths can be selected, and the entire process from the measurement to the calculation of Ca concentration is automated. Reaction process can be simultaneously monitored at multiple wavelengths.

■ Filter set (650-0157)

Contains the following filters:

Corning 9863. Band pass filter from 250 to 390nm only UV-29, UV-31, Cut off filters for wavelengths shorter UV-35, UV-39, than 290, 310, 350, 390 and 430nm UV-43

respectively





Long life xenon lamp (150W) (250-1600)

Performance guaranteed life: 500 hours (150 hours in case of standard lamp)



Solid sample holder (650-0161)

Optimized for the measurement of solid samples, powder samples, or highly concentrated solutions. It is designed to prevent the specular reflection from the sample surface from entering the emission monochromator. Includes a powder cell.

Sample thickness is 13mm max.



multiplier).



(5J0-0135)

(5.10-0136)

longer wavelengths.

Emission side 200 to 800nm correction range (200 to 600nm with standard light source) (Requires Photomultiplier R928F (650-1246).)



(650-0155)

(650-0156)

Used to measure the polarization angle in the UV-VIS region (with 650-0155) and in the VIS region (with 650-0156). The 650-0156 provides a higher accuracy in VIS region.

Wavelength range 260 to 700nm (650-0155) 380 to 730nm (650-0156)



Photomultiplier R928F (650-1246)

Enables a fluorescence measuring range of 200 to 900nm (200 to 700nm with standard photo-



■ Sub standard light source (115V)

Sub standard light source (220-240V)

Required for correction of emission spectrum at



Polarization accessory for UV/VIS

Polarization accessory for VIS

Automatic Polarization accessory for UV/VIS (5J0-0137)

Automatic Polarization accessory for VIS (5J0-0138)

Used for the measurement, calculation and data recording of fluorescence polarization angle and fluorescence anisotropy. Optimized for the measurement of antigen-anti body reaction, biological cells, proteins, enzymes, and other samples for the medical and biochemical fields.

Wavelength range	380 - 730nm (5J0-0137) 260 - 700nm (5J0-0138)
Polarizer rotation	0 to 90° automatic repetitive rotation on both excitation and emission sides
Measured items	Change of fluorescence polarization angle vs. time, fluorescence polarization angle, fluorescence anisotropy



Micro cell (650-0116)



■ Low scatter micro cell (650-0171)

Used for the measurement of trace samples of about 0.2mL with almost the same sensitivity as that obtained by using a 10mm cell.

The low scatter micro cell using a black guartz mask has a low scatter beam and is effective for high sensitivity analysis of trace samples.



Report generator program (5J0-0306)

Used to customize measurement reports. In addition to allowing user selection of size and position of report items, comments font, and graphs, calculations could be automatically executed using the spreadsheet function.

SPECIFICATIONS

ITEM	DESCRIPTION
Sensitivity	S/N 800 or better (RMS) using Raman band of water S/N 250 or better (Peak to Peak) Excitation wavelength 350nm, bandpass 5nm, response 2s
Minimum sample volume	0.6mL (in use of standard 10mm rectangular cell)
Photometric principle	Monochromatic light monitoring ratio calculation
Light source	150W xenon lamp, self-deozonating lamp house
Monochromator	Stigmatic concave diffraction grating: 900 lines/mm, F2.2 Brazed wavelength: Excitation side 300nm, emission side 400nm
Measuring wavelength range (on both EX and EM)	200 to 750nm, and zero-order light (Expandable up to 900nm with optional detector)
Bandpass	Excitation side: 1, 2.5, 5, 10, 20nm Emission side: 1, 2.5, 5, 10, 20nm
Resolution	1.0nm
Wavelength accuracy	1nm
Wavelength scan speed	30, 60, 240, 1,200, 2,400, 12,000, 30,000, 60,000nm/min
Wavelength drive speed	60,000nm/min
Response	Response from 0 to 98%: 0.002, 0.004, 0.01, 0.05, 0.1, 0.5, 2, 4 s
Photometric value range	-9999 to 9999
Data processing unit	PC: Windows [®] XP Professional
Printer	Printer compatible with Windows [®] XP
Dimensions/weight	Spectrophotometer: $620 \text{ W} \times 520 \text{ D} \times 300 \text{ H} \text{ mm}$ (excluding protrusions)/41kg
Working temperature /humidity	15 to 35°C, 45 to 80% (condensation not allowed, 70% or less at 35°C or higher)
Power consumption (spectrophotometer)	100, 115, 220, 230, 240 V AC, 50/60 Hz, 380 VA
FL Solutions program	Standard software

* Microsoft[®], Windows[®], Microsoft Excel, Microsoft Word and Windows[®] XP are registered trademarks, and other company names and merchandise names are registered trademarks or brand names of respective companies.

• FUNCTIONS

ITEM	DESCRIPTION
	Contour plotting (fluorescence/phosphorescence),
	bird's eye view
3-dimensional	Readout of EX/EM spectra from contour
measurement	Peak detection
	Calculation between files $(+, -, \times, \div)$
	Fluorescence/phosphorescence/luminescence spectra
	Synchronous spectra/repetitive measurement/CAT
	Excitation spectrum correction (200 to 600nm)
	Emission spectrum correction (200 to 600nm)
Wavelength scan	Excitation longer wavelength spectrum correction (500 to 800nm)
	Emission longer wavelength spectrum correction (500 to 800nm)
	Note: Sub standard light source (ontion) is necessary
	Tracing scale conversion graph axis conversion
	Smoothing
	Calculation between files $(+ - \times \pm)$
	Differentiation (first to fourth order)
	Contour plotting (fluorescence/phosphorescence)
	bird's eve view
3-dimensional time	Readout of time scan/EM spectra from contour
scan measurement	Peak detection
	Calculation between files $(+, -, \times, \div)$
	Time scan fluorescence/phosphorescence meas-
	urement mode (minimum data interval 1.0ms)
	Phosphorescence attenuation curve measurement
Time scan	Rate calculation
measurement	Tracing, scale conversion, graph axis conversion
mode	Smoothing
	Calculation between files $(+, -, \times, \div)$
	Differentiation (first to fourth order)
	Area calculation
Photometry mode	Quantitative analysis
	(fluorescence/phosphorescence/luminescence)
	Two/three-wavelength calculation
	Calibration curve (linear, quadratic, cubic,
	polygonal), factor enterable
	Peak ratio, peak area, quantization via differentiation
	Interruption, sample blank measurement, data deletion
	Calibration curve data correction, calibration
	Cumulative data averaging
	Statistic calculation
	Automatic constituity measurement function
Others	Pre-scan
	Data transport and graph conving to Microsoft® Evcel
	Print preview function

NOTICE: For proper operation, follow the instruction manual when using the instrument.

Specifications in this catalog are subject to change with or without notice, as Hitachi High-Technologies Corporation continues to develop the latest technologies and products for our customers.

Hitachi High-Technologies Corporation

Tokyo, Japan http://www.hitachi-hitec.com

24-14 Nishi-Shimbashi 1-chome, Minato-ku, Tokyo, 105-8717, Japan Tel: +81-3-3504-7211 Fax: +81-3-3504-7302

PRINTED WITH SOVINK, Printed with soy ink on 100% recycled paper. For further information, please contact your nearest sales representative.